



കേരള സർക്കാർ
Government of Kerala
2015



Regn. No. KERBIL/2012/45073
dated 5-9-2012 with RNI

Reg. No. KL/ TV(N)/634/2015-17

tl cf Kk ddv KERALA GAZETTE

ആധികാരികമായി പ്രസിദ്ധപ്പെടുത്തുന്നത്
PUBLISHED BY AUTHORITY

വാല്യം 4	തിരുവനന്തപുരം, ചൊവ്വ	2015 മാർച്ച് 3 3rd March 2015	നമ്പർ
Vol. IV	Thiruvananthapuram, Tuesday	1190 കുംഭം 19 19th Kumbham 1190	9
		1936 ഫാൽഗുനം 12 12th Phalguna 1936	No.

PART III

Stores Purchase

Medical Education Department

പുനർ ദർഘാസ് പരസ്യം

നമ്പർ സി8-18708 (1)/2014/ജി.എം.സി.റ്റി. 2015 ഫെബ്രുവരി 11.

തിരുവനന്തപുരം എസ്. എ. റ്റി. ആശുപത്രിയിലെ ഒബ്സ്റ്റട്രിക്സ് ആന്റ് ഗൈനക്കോളജി വിഭാഗത്തിലേക്കായി ആവശ്യമായ താഴെ പറയുന്ന സാധനം വിതരണം ചെയ്യുന്നതിന് മുദ്രവച്ച ദർഘാസുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു. ദർഘാസുകൾ ദർഘാസ് നമ്പരോടുകൂടി നിശ്ചിത തീയതിക്ക് മുൻപ് പ്രിൻസിപ്പാൾ, ഗവൺമെന്റ് മെഡിക്കൽ കോളേജ്, തിരുവനന്തപുരം - 695 011 എന്ന വിലാസത്തിൽ എത്തിച്ചിരിക്കേണ്ടതാണ്.

ദർഘാസ് നമ്പരും സാധനത്തിന്റെ പേരും എണ്ണവും—
185/2014—സർജിക്കൽ ഡയറക്ടർ-1 എണ്ണം.

ദർഘാസ് ഫോറത്തിന്റെ വില :

അസ്സൽ— 2,520.

ശരിപ്പകർപ്പ്— 1,260.

ദർഘാസ് ഫോറം വിൽക്കുന്ന അവസാന തീയതിയും സമയവും—5-3-2015, 3.30 മണി.

ദർഘാസ് ഫോറം സ്വീകരിക്കുന്ന അവസാന തീയതിയും സമയവും—6-3-2015, 2.30 മണി.

ദർഘാസ് തുറക്കുന്ന തീയതിയും സമയവും—6-3-2015, 3 മണി.

ദർഘാസ് ഫോറങ്ങൾ തപാലിൽ വേണ്ടവർ അയയ്ക്കുന്നതിനുവേണ്ട തുക (` 50) അധികമായി അയയ്ക്കേണ്ടതാണ്. ദർഘാസ് ഫോറത്തിന്റെ വില ഓഫീസിൽ നേരിട്ടോ മണിയോർഡർ ആയോ സ്വീകരിക്കുന്നതാണ്. അല്ലാത്തതുള്ള എല്ലാ അപേക്ഷകളും നിരസിക്കുന്നതാണ്. ഫോറത്തോടൊപ്പം മുദ്രപത്രങ്ങളോ മറ്റേതെങ്കിലും ഡോക്യുമെന്റുകളോ അയയ്ക്കാൻ ആവശ്യപ്പെട്ടുകൊണ്ടുള്ളതും മേൽവിലാസം വ്യക്തമല്ലാത്തതുമായ മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. മണിയോർഡറുകൾവഴി ഫോറങ്ങൾ ആവശ്യപ്പെട്ടുകൊണ്ടുള്ള അപേക്ഷകൾ ഈ ഓഫീസിൽ 27-2-2015-ന് മുമ്പ് കിട്ടിയിരിക്കണം. താമസിച്ചുകിട്ടുന്ന മണിയോർഡറുകൾ നിരസിക്കുന്നതാണ്. ദർഘാസ് ഫോറം കൈമാറാൻ പാടില്ലാത്തതും ഒരിക്കൽ അടച്ച തുക തിരിച്ചു കിട്ടുന്നതുമല്ല. താമസിച്ചു കിട്ടുന്ന ദർഘാസുകൾ യാതൊരു കാരണവശാലും സ്വീകരിക്കുന്നതല്ല. കൂടുതൽ വിവരങ്ങൾ പ്രവൃത്തി ദിവസങ്ങളിൽ ഈ ഓഫീസിൽനിന്നും ലഭിക്കുന്നതാണ്.

No. C8/18708 (1) /2014/GMCT. 10th February 2015.

Sealed competitive tenders in the prescribed performa are invited for the purchase of the following item required for the Professor & Head, Obstetrics and Gynaecology Department, SAT Hospital, Thiruvananthapuram. Tender should be superscribed tender number and addressed to the Principal, Government Medical College, Thiruvananthapuram, Pin-695 011. Details can be had from this office on working hours till 3.30 p. m.

Tender No. and Name of item with Quantity—185/2014—Surgical Diathermy—1 No.

Department—Obstetrics & Gynaecology.

Cost of tender form (inclusive of VAT) :

Original—` 2,520.

Duplicate—` 1,260.

Last date and time of sale of tender form—5-3-2015, 3.30 p. m.

Last date and time of receipt of tender form—6-3-2015, 2.30 p. m.

Opening date and time of tender—6-3-2015, 3 p. m.

Tender form will be sent through post if required provided due amount with sufficient postage charge (` 50) is sent by way of Money Order in favour of The Principal, Government Medical College, Thiruvananthapuram, Kerala State, Pin-695 011. Money Order should reach this office on or before 27-2-2015. No other forms of remittance will be accepted.

Specification for Electrosurgical Generator/Cautery

1. Micro controller based isolated Electrosurgical Generator having both Monopolar and Bipolar outputs designed for all surgical procedures.
2. Generator should be able to monitor changes in tissue impedance continuously and adjusts power, power efficiency rating of 98% or more for accurate and consistent cut performance.
3. Monopolar outputs should have three cutting modes:—

- (a) Low Cut for delicate tissue or Laparoscopic cases having maximum power of 300 W.
- (b) Pure cut for clean, precise cut in general surgery having maximum power of 300 W.
- (c) Blend mode for cutting with homeostasis having maximum power of 200 W.

All cut modes should be able to adjust output power depending on tissue density by less than 15% or 5W, whichever is greater.

4. It should have Four monopolar Coagulation Modes with maximum power of 120 W
 - (a) Desiccate mode for low voltage contact coagulation suitable for endoscopic and delicate tissue work.
 - (b) Fulgurate mode for efficient non-contact coagulation in most applications.
 - (c) Fulguration mode for lower voltage coagulation (low crest factor)
 - (d) Spray mode should have randomized spray effect of varying amplitude and frequency for coagulating large tissue areas with minimum depth of necrosis.

5. It should have three bipolar modes with maximum power of 70 W.
 - (a) Precise mode to have fine control of desiccation in delicate tissue.
 - (b) Standard mode for applications at low voltage to prevent sparking.
 - (c) Macro mode for applications on tissue with high resistance.
6. It should have patient plate monitoring facility and should give audiovisual alarm and deactivate output if contact between patient and patient plate is not proper to eliminate the risk of patient burns.
7. The system should be able to connect two monopolar foot switches simultaneously and can use two monopolar ports simultaneously.
8. The unit should have two handswitching and two footswitching Monopolar outputs and one handswitching and footswitching bipolar output.
9. It should have membrane keyboard for power settings.
10. The unit should have individual digital display of power for Bipolar, Monopolar cut and Monopolar Coag.
11. The unit should have temperature sensing cooling fan, which should operate automatically to protect generator from thermal damage.
12. It should have RS232 serial port to interface with computer.
13. The machine should facilitate under water cutting with coagulating
14. It should be compatible with argon plasma coagulator system and CUSA.
15. The unit should have RF activation port to tell other equipments like ECG or EEG that RF current is being generated.
16. The unit should not have RF leakage current more than 150mA.
17. The unit should be operational between 170V AC, 50Hz. (should not require stabilizer)
18. Should be type CF Equipment (IEC 601-1 & IEC 601-2-2) Defibrillator Proof.
19. The system and all its accessories should be USFDA certified.
20. Accessories

- Monopolar Footswitch—1 No.
- Bipolar footswitch —1 No.
- Disposable Patient Plate Monitoring System— 100 Nos.
- Bipolar Forceps—3 Nos.
- Universal Adapter—1 No.
- Disposable Hand switching pencil—100 Nos.

TENDER SCHEDULE

1. The period for which the rates to be in firm One year.
2. The rates should be inclusive of all taxes, otherwise tender will be rejected without any notice and free delivery at the department concerned.
3. Advance payments conditions will not be considered.
4. Tender forms are neither transferable nor the cost of it will be refundable under any circumstances.
5. Tenders are to be sent with details and illustrated pamphlet (wherever necessary) neatly typed. Manuscripts should be avoided as far as possible.
6. Tenders should be accompanied with a formal agreement executed in ` 100 Kerala non-judicial-Stamp Paper together with required sum of Earnest Money Deposits.
7. Earnest Money Deposits (One Percentage) of the total cost of the articles quoted for, can be remitted by way of crossed Demand Drafts drawn in favour of the Principal, Government Medical College, Thiruvananthapuram.
8. Firms will be exempted from furnishing Earnest Money Deposit if they had produced the duly attested copies of relevant certificate along with their tenders.
9. Tenders unsealed, unsigned or submitted without superscription as specified are liable to be rejected.
10. Tenders submitted with insufficient Earnest Money Deposit will not be entertained.

11. Belated tenders will not be considered under any circumstances.
12. For Medicines, Surgicals and Materials, samples should be submitted along with the tender. Tenders without samples will not be considered for selection.
13. Tenders should bear the signature and full postal address of the tenderers and witness wherever necessary.
14. The undersigned reserves the right to reject any tender wholly or partially without any prior information to the tenders.
15. Specifications of the items are given in the list.
16. Negotiation will be made only with the firm offering lowest price whose specifications confirms to the requirements in the tender notice.
17. Tender will be opened in the presence of each of the tenderers or his authorized representatives who may present at the time.
18. More details can be had from this office on working time.

Govt. Medical College,
Thiruvananthapuram.

(Sd.)
For Principal.

ദർഘാസ് പരസ്യം

നമ്പർ എസ്1/16978/2014.

2015 ജനുവരി 12.

കോഴിക്കോട് മെഡിക്കൽ കോളേജ് ആശുപത്രിയിലേക്ക് താഴെ പറയുന്നവ സപ്ലൈ ചെയ്യുന്നതിന് നിശ്ചിത ഫോറത്തിൽ സീൽ ചെയ്ത ടെണ്ടറുകൾ ക്ഷണിച്ചുകൊള്ളുന്നു.

ക്രമ നമ്പർ	സാധന വിവരം	അളവ്
1	ബ്ലഡ് കലക്ഷൻ ട്യൂബ് (നോൺ വാക്വം ട്യൂബ് വിത്ത് ക്ലോട്ട് ആക്റ്റിവേറ്റർ)	300000
2	ബ്ലഡ് കലക്ഷൻ ട്യൂബ് ഫോർ ഗ്ലൂക്കോസ് (ഫ്ലൂറൈഡ് ഓക്സലേറ്റ് മിക്സ്ചർ കണ്ടെയ്നിംഗ് ട്യൂബ്)	200000

അടങ്കൽ തുക—` 15,00,000.

നിരതദ്രവ്യം—1%.

ടെണ്ടർ അയയ്ക്കുന്ന കവറിനുമുകളിൽ നമ്പർ എസ്1/16978/2014 എന്ന് വ്യക്തമായി കാണിച്ചിരിക്കേണ്ടതാണ്.

ടെണ്ടർ ഫോറത്തിന്റെ വില—` 2,634.

ടെണ്ടർ ഫോറം വിൽക്കുന്ന അവസാന തീയതിയും സമയവും—2-3-2015, 3 മണി.

ടെണ്ടർ കിട്ടേണ്ട അവസാന തീയതിയും സമയവും—3-3-2015, 5 മണി.

ടെണ്ടർ തുറക്കുന്ന തീയതിയും സമയവും—4-3-2015, 12 മണി.

ടെണ്ടറിൽ പറയുന്ന നിരക്കുകൾ പ്രാബല്യത്തിലിരിക്കേണ്ട അവസാന തീയതി—30-6-2015.

ടെണ്ടർ ഫോറം കിട്ടുന്നതിനും പുരിപ്പിച്ചു അയയ്ക്കേണ്ട തുമായ വിലാസം—സൂപ്രണ്ട്, മെഡിക്കൽ കോളേജ് ആശുപത്രി, കോഴിക്കോട്.

No. S1/16978/2014.

12th January 2015.

Sealed tenders in the prescribed form are invited for the supply of the following items to Medical College Hospital, Kozhikode.

Sl. No.	Item	Quantity
1	Blood Collection Tube (Non Vacuum Tube with Clot activator)	300000
2	Blood Collection Tube for Glucose (Fluoride Oxalate Mixture containing tube)	200000

PAC—` 15,00,000.

EMD—1%.

Superscription tender No. should be specifically noted as S1/16978/2014.

Price of tender (in duplicate) ST Cash only/Money order—` 2,634.

Date and time of close of sale of tender forms—2-3-2015, 3 p.m.

Date and time for receipt of tender—3-3-2015, 5 p. m.

Date and time for opening of tender—4-3-2015, 12 Noon.

Date upto which rates are to be firms—30-6-2015.

Address of Officer from whom tender forms are to be obtained and whom tenders are to be sent— The Superintendent, Medical College Hospital, Kozhikode.

Medical College Hospital,
Kozhikode.

(Sd.)
Superintendent.